

Product data sheet

UHV gate valve, Series 108, DN 250 (ID 10") Ordering No. 10848-TE28

Description

Flange ASA-LP 250

Actuator pneumatic double acting, with 3-position actuator

with position indicator

Feedthrough Bellows

Technical data

Leak rate - Valve body $< 5 \cdot 10^{-10}$ mbar Is⁻¹

- Valve seat $< 1 \cdot 10^{-9}$ mbar Is⁻¹

Pressure range $1 \cdot 10^{-10}$ mbar to 1.2 bar (abs)

Differential pressure on the gate ≤ 1.2 bar

Differential pressure at opening ≤ 30 mbar

Conductance (molecular flow) $\leq 20000 \text{ ls}^{-1}$

Cycles until first service 50 000 (unheated and under clean conditions)

Temperature – Valve body ≤ 250 °C open / ≤ 200 °C closed (bake-out max. 24h)

(Maximum values: depending — Actuator \leq 200 °C on operating conditions and — Position indicator \leq 80 °C

sealing materials)

Heating and cooling rate 50 °C h⁻¹

Material (main components) – Valve body AISI 304 (1.4301)

- Mechanism AISI 316L (1.4404), AISI 304 (1.4301)

- Bellows AISI 316L (1.4404, 1.4435)

Seal – Bonnet metal

Gate
 Actuator
 FKM (Viton[®]), O-ring
 FKM (Viton[®]), NBR

Mounting position any

Volume of pneumatic actuator 0.35 I / 0.0122 ft³

Compressed air 5-7 bar/73-102 psi

min. - max. overpressure

Compressed air connection 1/8" ISO / NPT

Actuation time – closing 4 s

opening4 s

Weight 55 kg / 120 lbs

Created by: MAEM	Release date: 2015-02-24	1 of 2
Modified by:	Release date:	779369EA



Product data sheet

UHV gate valve, Series 108, DN 250 (ID 10") Ordering No. 10848-TE28

Behavior in case of compressed

air pressure drop

- Valve closed - Valve open

valve remains closed undefined

Middle position

undefined

Behavior in case of power failure

 Valve closed - Valve open

depending on customer installation depending on customer installation


Middle position

depending on customer installation

Position indicator

Type Micro switch Voltage ≤ 50 V AC / DC

Current max. ≤ 1.2 A



Wiring diagram

Created by: MAEM	Release date: 2015-02-24	2 of 2
Modified by:	Release date:	779369EA