

Product data sheet

HV gate valve, Series 140, DN 160 (6") Ordering No. 14044-CE28

Description

Flange CF-F 160

Actuator Pneumatic, with 3-position actuator

with position indicator

Feedthrough Rotary feedthrough

Technical data

Leak rate - Valve body $< 1 \cdot 10^{-9}$ mbar ls⁻¹

– Valve seat < 1 · 10⁻⁰ mbar ls⁻¹</p>

Pressure range $1 \cdot 10^{-8}$ mbar to 2 bar (abs)

Differential pressure on the gate ≤ 2 bar

Differential pressure at opening \leq 30 mbar

Conductance (molecular flow) – Nominal 5'150 ls⁻¹

– Min. adjustable
– Unheated and under
200'000

Unheated and under

clean conditions

Temperature – Valve body ≤ 150 °C (Maximum values: depending – Actuator ≤ 50 °C

(Maximum values: depending — Actuator ≤ 50 °C on operating conditions and — Position indicator ≤ 80 °C

sealing materials)

Cycles until first service

Heating and cooling rate 50 °C h⁻¹

Material (main components) – Valve body AISI 304 (1.4301)

– Mechanism
 AISI 304 (1.4301), AISI 301 (1.4310),

AISI 316L (1.4404)

Seal – Bonnet FKM (Viton®), vulcanized

– Gate
 – Actuator
 FKM (Viton®), O-ring
 FKM (Viton®), NBR

Mounting position any

Volume of pneumatic actuator 0.96 I / 0.034 ft³

Compressed air 4-7 bar / 58-102 psi

min. - max. overpressure

Compressed air connection G\(^1\)8" (\(^1\)8" NPT for USA)

Actuation time - closing $\leq 3.5 \text{ s}$

- opening ≤ 3.5 s

Weight 30 kg / 66 lbs

Created by: NIW	Release date: 27.10.2020	1/2
Modified by:	Release date:	1061112EA



Product data sheet

HV gate valve, Series 140, DN 160 (6") Ordering No. 14044-CE28

Behavior in case of compressed

- Valve closed - Valve open

valve remains closed

air pressure drop

- Middle position

undefined undefined

Behavior in case of power failure

- Valve closed

depending on customer installation depending on customer installation

- Valve open - Middle position

depending on customer installation

Valve position indication

visual (mechanical)

Electrical connections

Position indicator

Type Micro switch

Voltage ≤ 250 V AC ≤ 50 V DC

Current max. ≤ 5 A ≤ 3 A

2		OPEN				
_4	1 INTERMEDIATE					
<u>6</u> 5	6 5 CLOSED					
Valve/Ventil/Vanne						
		OPEN	INTERMEDIATE	CLOSED		
(3)	1,2	connected	4	=		
	4,3	I	connected	-		
Front view M-2770-008	6,5	Œ	8	connected		

Wiring diagram

Created by: NIW	Release date: 27.10.2020	2/2
Modified by:	Release date:	1061112EA