

Product data sheet

HV gate valve, Series 140, DN 160 (6") Ordering No. 14044-JE28

Description

Flange JIS 160

Actuator Pneumatic, with 3-position actuator

with position indicator

Feedthrough Rotary feedthrough

Technical data

Leak rate - Valve body $< 1 \cdot 10^{-9}$ mbar ls⁻¹

– Valve seat < 1 · 10⁻⁰ mbar ls⁻¹</p>

Pressure range $1 \cdot 10^{-8}$ mbar to 2 bar (abs)

Differential pressure on the gate ≤ 2 bar

Differential pressure at opening \leq 30 mbar

Conductance (molecular flow) – Nominal 5'150 ls⁻¹

– Min. adjustable
7.5 ls⁻¹
Cycles until first service
– Unheated and under
200'000

clean conditions

Temperature – Valve body ≤ 150 °C (Maximum values: depending – Actuator ≤ 50 °C

on operating conditions and — Position indicator

sealing materials)

Heating and cooling rate 50 °C h⁻¹

Material (main components) – Valve body AISI 304 (1.4301)

– Mechanism AISI 304 (1.4301), AISI 301 (1.4310),

AISI 316L (1.4404)

≤ 80 °C

Seal – Bonnet FKM (Viton®), vulcanized

– Gate
– Actuator
FKM (Viton®), O-ring
FKM (Viton®), NBR

Mounting position any

Volume of pneumatic actuator 0.96 I / 0.034 ft³

Compressed air 4-7 bar / 58-102 psi

min. - max. overpressure

Compressed air connection G\(^1\)8" (\(^1\)8" NPT for USA)

Actuation time - closing $\leq 3.5 \text{ s}$

- opening ≤ 3.5 s

Weight 30 kg / 66 lbs

Created by: NIW	Release date: 27.10.2020	1/2
Modified by:	Release date:	1061116EA



Product data sheet

HV gate valve, Series 140, DN 160 (6") Ordering No. 14044-JE28

Behavior in case of compressed

sed - Valve closed

valve remains closed

air pressure drop

Valve openMiddle position

undefined undefined

Behavior in case of power failure

Valve closed

depending on customer installation depending on customer installation

Valve openMiddle position

depending on customer installation

Valve position indication

visual (mechanical)


Electrical connections

Position indicator

Type Micro switch

Voltage \leq 250 V AC \leq 50 V DC

Current max. $\leq 5 \text{ A} \leq 3 \text{ A}$



Wiring diagram

Created by: NIW	Release date: 27.10.2020	2/2
Modified by:	Release date:	1061116EA