



VAT Vakuumventile AG  
CH-9469 Haag, Schweiz

## Product data sheet

**Vacuum angle valve**  
**Series 244, DN 100 (ID 4'')**  
**Ordering No. 24440-QA11-0001**

### Description

Flange	ISO-K 100
Actuator	Pneumatic, single acting with closing spring – without solenoid valve – without position indicator
Feedthrough	Shaft feedthrough

### Technical data

Leak rate (static)	– Valve body – Valve seat	$< 1 \cdot 10^{-9}$ mbar ls <sup>-1</sup> $< 1 \cdot 10^{-9}$ mbar ls <sup>-1</sup>
Pressure range		$1 \cdot 10^{-7}$ mbar to 2 bar (abs)
Differential pressure on the plate	– In opening direction – In closing direction	$\leq 1.2$ bar $\leq 2$ bar
Differential pressure at opening		$\leq 1$ bar
Conductance (molecular flow)		440 ls <sup>-1</sup>
Actuation time	– Closing – Opening	$\leq 1$ sec. $\leq 1$ sec.
Cycles until first service (Tmax 80°C, under clean conditions)		1 million
Temperature (Maximum values: depending on operating conditions and sealing materials)	– Valve body – Actuator	$\leq 150$ °C $\leq 120$ °C
Material	– Valve body – Actuator – Plate – Shaft feedthrough – Spring – Snap ring for axle	Aluminum (EN AC-42000) Aluminum Aluminum (EN AW 6082) Aluminum (EN AW 6082) AISI 301 (1.4310) AISI 301 (1.4310)
Seal	– Bonnet O-ring – Plate O-ring – Feedthrough O-ring – Actuator O-ring	FKM (Viton®) FKM (Viton®) FKM (Viton®) FKM (Viton®)
Weight		9 kg / 20 lbs
Mounting position		any
Volume of pneumatic actuator		0.3 l / 0.01 ft <sup>3</sup>
Compressed air min. – max. overpressure		4.5 – 7 bar / 65 – 102 psi
Compressed air connection		1/8" ISO/NPT female
Behavior in case of compressed air pressure drop	– Valve closed – Valve opened – During actuation	Valve remains closed and leaktight Valve closes leaktight Valve closes leaktight

Created by: DIC	Release date: 15.04.2016	1/2
Modified by:	Release date:	<b>600101EC</b>



VAT Vakuumventile AG  
CH-9469 Haag, Schweiz

## Product data sheet

**Vacuum angle valve**

**Series 244, DN 100 (ID 4'')**

**Ordering No. 24440-QA11-0001**

Behavior in case of power failure	- Valve closed	Valve remains closed and leaktight
	- Valve opened	Depending on customer installation
	- During actuation	Depending on customer installation

Created by: DIC	Release date: 15.04.2016	2/2
Modified by:	Release date:	<b>600101EC</b>