

## Product data sheet Vacuum angle valve Series 244, DN 160 (ID 6'') Ordering No. 24444-QE14-0001

## Description

Flange		ISO-K 160
Actuator		Pneumatic, double acting – without solenoid valve – without position indicator
Feedthrough		Shaft feedthrough
Technical data		
Leak rate (static)	– Valve body – Valve seat	< 1 · 10 <sup>-9</sup> mbar Is <sup>-1</sup> < 1 · 10 <sup>-9</sup> mbar Is <sup>-1</sup>
Pressure range		$1 \cdot 10^{-7}$ mbar to 2 bar (abs)
Differential pressure on the plate	<ul> <li>In opening direction</li> <li>In closing direction</li> </ul>	≤ 1.2 bar ≤ 2 bar
Differential pressure at opening		$\leq$ 1 bar [1 bar with reduced number of cycles]
Conductance (molecular flow)		1000 ls <sup>-1</sup>
Actuation time	– Closing – Opening	≤ 1.5 sec. ≤ 1.5 sec.
Cycles until first service (Tmax 80°C, under clean conditions		2 million
Temperature (Maximum values: depending on operating conditions and sealing materials)	<ul> <li>Valve body</li> <li>Actuator</li> </ul>	≤ 150 °C ≤ 120 °C
Material	<ul> <li>Valve body</li> <li>Actuator</li> <li>Plate</li> <li>Shaft feedthrough</li> </ul>	AISI 316L (1.4404) Aluminum AISI 316L (1.4404) AISI 304 (1.4301)
Seal	<ul> <li>Bonnet O-ring</li> <li>Plate O-ring</li> <li>Feedthrough O-ring</li> <li>Actuator O-ring</li> </ul>	FKM (Viton®) FKM (Viton®) FKM (Viton®) FKM (Viton®)
Weight		15 kg / 33 lbs
Mounting position		any
Volume of pneumatic actuator		0.35 I / 0.012 ft <sup>3</sup>
Compressed air min. – max. overpressure		4.5 – 7 bar / 65 – 102 psi
Compressed air connection		1∕₃" ISO/NPT female
Behavior in case of compressed air pressure drop	<ul> <li>Valve closed</li> <li>Valve opened</li> <li>During actuation</li> </ul>	depending on customer installation depending on customer installation depending on customer installation

Created by: LIN	Release date: 19.11.2013	1/2
Modified by:	Release date:	604913EA



## Product data sheet

Vacuum angle valve Series 244, DN 160 (ID 6'') Ordering No. 24444-QE14-0001

Behavior in case of power failure - Valve closed

- Valve closedValve opened
- During actuation

depending on customer installation depending on customer installation depending on customer installation

Created by: LIN	Release date: 19.11.2013	2/2
Modified by:	Release date:	604913EA