



VAT Vakuumventile AG
CH-9469 Haag, Schweiz

Product data sheet

HV angle valve

Series 264, DN 63 (ID 2 1/2")

Ordering No. 26436-QA11-0002

Description

Flange	ISO-K 63
Actuator	Pneumatic, single acting with closing spring – without solenoid valve – without position indicator
Feedthrough	Bellows

Technical data

Leak rate	– Valve body – Valve seat	$< 1 \cdot 10^{-9}$ mbar ls ⁻¹ $< 1 \cdot 10^{-9}$ mbar ls ⁻¹
Pressure range		$1 \cdot 10^{-8}$ mbar to 4 bar (abs)
Differential pressure on the plate	– In opening direction – In closing direction	≤ 1.2 bar ≤ 4 bar
Differential pressure at opening		≤ 1 bar
Conductance (molecular flow)		160 ls ⁻¹
Closing time		≤ 0.7 sec.
Cycles until first service (Tmax 80°C, under clean conditions)		3 million
Temperature (Maximum values: depending on operating conditions and sealing materials)	– Valve body – Actuator	≤ 150 °C ≤ 120 °C
Material	– Valve body – Actuator – Plate – Bellows	Aluminum (EN AW-6060) Aluminum AISI 316L (1.4435 ESU) AISI 316L (1.4404, 1.4435), AISI 316 Ti (1.4571)
Seal	– Bonnet O-ring – Plate O-ring – Actuator O-ring	FKM (Viton®) FKM (Viton®) FKM (Viton®)
Weight		3.6kg / 7.936 lbs
Mounting position		any
Volume of pneumatic actuator		0.112 l / 0.0040 ft ³
Compressed air min. – max. overpressure		4 – 8 bar / 60 – 115 psi
Compressed air connection		Female 1/8" (ISO/NPT)

Created by: WARO	Release date: 10.06.2014	1/2
Modified by:	Release date:	753420EA



VAT Vakuumventile AG
CH-9469 Haag, Schweiz

Product data sheet

HV angle valve

Series 264, DN 63 (ID 2 1/2")

Ordering No. 26436-QA11-0002

Behavior in case of compressed air pressure drop	<ul style="list-style-type: none">- Valve closed- Valve opened- During actuation	<ul style="list-style-type: none">valve remains closed and leaktightvalve closes leaktightvalve closes leaktight
Behavior in case of power failure	<ul style="list-style-type: none">- Valve closed- Valve opened- During actuation	<ul style="list-style-type: none">valve remains closed and leaktightdepending on customer installationdepending on customer installation

Created by: WARO	Release date: 10.06.2014	2/2
Modified by:	Release date:	753420EA