



VAT Vakuumventile AG  
CH-9469 Haag, Schweiz

## Product data sheet

### HV inline valve

Series 265, DN 25 (ID 1")

Ordering No. 26528-KE11-0003

## Description

Flange	ISO-KF 25
Actuator	Pneumatic, single acting with closing spring – without solenoid valve – without position indicator
Feedthrough	Bellows

## Technical data

Leak rate	– Valve body – Valve seat	$< 1 \cdot 10^{-9}$ mbar ls <sup>-1</sup> $< 1 \cdot 10^{-9}$ mbar ls <sup>-1</sup>
Pressure range		$1 \cdot 10^{-8}$ mbar to 5 bar (abs)
Differential pressure on the plate	– In opening direction – In closing direction	$\leq 2$ bar $\leq 5$ bar
Differential pressure at opening		$\leq 1$ bar
Conductance (molecular flow)		14 ls <sup>-1</sup>
Closing time		$\leq 0.2$ sec.
Cycles until first service (T <sub>max</sub> 80 °C, under clean conditions)		3 million
Temperature (Maximum values: depending on operating conditions and sealing materials)	– Valve body – Actuator	$\leq 150$ °C $\leq 120$ °C
Material	– Valve body – Actuator – Plate – Bellows	AISI 304 (1.4301) Aluminum AISI 316L (1.4404) AISI 316L (1.4404, 1.4435) AISI 316 Ti (1.4571)
Seal	– Bonnet O-ring – Plate O-ring – Actuator O-ring	FKM (Viton®) FKM (Viton®) FKM (Viton®)
Weight		0.55 kg / 1.21 lbs
Mounting position		any
Volume of pneumatic actuator		0.011 l / 0.0004 ft <sup>3</sup>
Compressed air min. – max. overpressure		4 – 8 bar / 58 – 116 psi
Compressed air connection		M5 female (10-32 UNF suitable)

Created by: WARO	Release date: 07.05.2014	1/2
Modified by:	Release date:	<b>748592EA</b>



VAT Vakuumventile AG  
CH-9469 Haag, Schweiz

## Product data sheet

**HV inline valve**

**Series 265, DN 25 (ID 1")**

**Ordering No. 26528-KE11-0003**

Behavior in case of compressed air pressure drop	<ul style="list-style-type: none"><li>- Valve closed</li><li>- Valve opened</li><li>- During actuation</li></ul>	<ul style="list-style-type: none"><li>valve remains closed and leaktight</li><li>valve closes leaktight</li><li>valve closes leaktight</li></ul>
Behavior in case of power failure	<ul style="list-style-type: none"><li>- Valve closed</li><li>- Valve opened</li><li>- During actuation</li></ul>	<ul style="list-style-type: none"><li>valve remains closed and leaktight</li><li>depending on customer installation</li><li>depending on customer installation</li></ul>

Created by: WARO	Release date: 07.05.2014	2/2
Modified by:	Release date:	<b>748592EA</b>