



VAT Vakuumventile AG
CH-9469 Haag, Schweiz

Product data sheet

HV inline valve

Series 265, DN 50 (ID 2")

Ordering No. 26534-KA11-0001

Description

Flange	ISO-KF 50
Actuator	Pneumatic, single acting with closing spring – without solenoid valve – without position indicator
Feedthrough	Bellows

Technical data

Leak rate	– Valve body – Valve seat	$< 1 \cdot 10^{-9}$ mbar ls ⁻¹ $< 1 \cdot 10^{-9}$ mbar ls ⁻¹
Pressure range		$1 \cdot 10^{-8}$ mbar to 5 bar (abs)
Differential pressure on the plate	– In opening direction – In closing direction	≤ 2 bar ≤ 5 bar
Differential pressure at opening		≤ 1 bar
Conductance (molecular flow)		80 ls ⁻¹
Closing time		≤ 0.65 sec.
Cycles until first service (T _{max} 80 °C, under clean conditions)		3 million
Temperature (Maximum values: depending on operating conditions and sealing materials)	– Valve body – Actuator	≤ 150 °C ≤ 120 °C
Material	– Valve body – Actuator – Plate – Bellows	Aluminum (EN AW-6082) Aluminum AISI 316L (1.4404) AISI 316L (1.4404, 1.4435) AISI 316 Ti (1.4571)
Seal	– Bonnet O-ring – Plate O-ring – Actuator O-ring	FKM (Viton®) FKM (Viton®) FKM (Viton®)
Weight		2.50 kg / 5.51 lbs
Mounting position		any
Volume of pneumatic actuator		0.047 l / 0.0017 ft ³
Compressed air min. – max. overpressure		4 – 8 bar / 58 – 116 psi
Compressed air connection		M5 female (10-32 UNF suitable)

Created by: WARO	Release date: 15.05.2014	1/2
Modified by:	Release date:	750343EA



VAT Vakuumventile AG
CH-9469 Haag, Schweiz

Product data sheet

HV inline valve

Series 265, DN 50 (ID 2")

Ordering No. 26534-KA11-0001

Behavior in case of compressed air pressure drop	<ul style="list-style-type: none">- Valve closed- Valve opened- During actuation	<ul style="list-style-type: none">valve remains closed and leaktightvalve closes leaktightvalve closes leaktight
Behavior in case of power failure	<ul style="list-style-type: none">- Valve closed- Valve opened- During actuation	<ul style="list-style-type: none">valve remains closed and leaktight depending on customer installationdepending on customer installation

Created by: WARO	Release date: 15.05.2014	2/2
Modified by:	Release date:	750343EA