

## **Product data sheet**

# UHV angle valve, Series 284, DN 25 (ID 1") Ordering No. 28428-GE21-CVK1

## **Description**

Flange CF-R 25

Actuator Pneumatic, single acting with closing spring

without solenoid valvewith position indicator

Feedthrough Bellows

Options CVK = - Product: valve type ITER 22

Actuator: seal material EPDMGate: seal material EPDM

- Position indicator: double; Glenair 4 pin connector;

120°C

- Compressed air connection: female 1/8" ISO/NPT

#### Technical data<sup>1</sup>

Leak rate - Valve body  $\leq 1 \cdot 10^{-8}$  mbar Is<sup>-1</sup>

– Valve seat < 1 · 10⁻¹ mbar ls⁻¹</p>

Leak rate (accidental condition) — Valve body  $\leq 1 \cdot 10^{-8}$  mbar Is<sup>-1</sup>

Valve seat n/a

Pressure range  $1 \cdot 10^{-8}$  mbar to 1.5 bar (abs)

Differential pressure on the plate - In either direction  $\leq 1.5$  bar Differential pressure at opening  $\leq 1$  bar Conductance (molecular flow) 14 ls<sup>-1</sup>

Closing time ≤ 0.2 sec.

Cycles until first service 10 000

(Tmax 80 °C, under clean conditions)

Temperature - Valve body  $\le 150$  °C (Maximum values: depending on operating conditions and sealing - Position indicator  $\le 150$  °C  $\le 120$  °C

Material – Valve body AISI 304 (1.4301)

ActuatorAluminum

Plate AISI 316L (1.4404)

- Bellows AISI 316L (1.4404, 1.4435), AISI 316 Ti (1.4571)

Seal – Bonnet metal – Plate EPDM

- Actuator EPDM

Weight 0.65 kg / 1.433 lbs

Mounting position any

Radiation resistance 10<sup>6</sup> Gy (10<sup>8</sup> rad)

Maximum magnetic field levels ≤ 150 mT

<sup>&</sup>lt;sup>1</sup> Values according technical specification D\_R8752R v.1.1

Created by: GARM	Release date: 13.02.2019	1/2
Modified by: Phil Schneider	Release date: 15.02.2021	957805EB



## **Product data sheet**

# UHV angle valve, Series 284, DN 25 (ID 1") Ordering No. 28428-GE21-CVK1

Volume of pneumatic actuator

Compressed air

min. - max. overpressure

Compressed air connection

Behavior in case of compressed

air pressure drop

- Valve closed

Valve open

During actuation

Behavior in case of power failure – Valve closed – Valve open

During actuation

- During actuation

Assembly cleanliness level

0.011 I / 0.0004 ft<sup>3</sup>

4 – 8 bar / 58 – 116 psi

1/8" female ISO/NPT

valve remains closed and leaktight

valve closes leaktight valve closes leaktight

depending on customer installation

depending on customer installation depending on customer installation

ISO Class 8

#### **Related documents**

Dimensional drawing No. STEP file No.

956348 956347

### **Electrical connections**

### Position indicator (2x)

Type Micro switch Voltage  $\leq 50 \text{ V AC / DC}$  Current 5 - 100 mA

Short circuit protected No

Connector on the valve: 8070-2530-02Z16-4GA

Mating connector:\* 8070-3039-01Z16-4KA

(pins compatible with AWG 20

wire)

2	OPEN			
CLOSED/INTERMEDIATE  Valve/Ventil/Vanne				
ļ		INTERMEDIATE		
1,2	connected	-	-	
1,4	-	connected	-	
Front view 1,3	-	-	connected	
Wiring diagram				

<sup>\*</sup>The mating connector is not part of the valve and needs to be ordered separately at Glenair.

Created by: GARM	Release date: 13.02.2019	2/2
Modified by: Phil Schneider	Release date: 15.02.2021	957805EB